



Cancer Research

Integrated Microfluidic and Imaging Platform for a Kinase Activity Radioassay to Analyze Minute Patient Cancer Samples

Cong Fang, Yanju Wang, Nam T. Vu, et al.

Cancer Res 2010;70:8299-8308. Published OnlineFirst September 13, 2010.

Updated Version	Access the most recent version of this article at: doi: 10.1158/0008-5472.CAN-10-0851
Supplementary Material	Access the most recent supplemental material at: http://cancerres.aacrjournals.org/content/suppl/2010/09/13/0008-5472.CAN-10-0851.DC1.html

Cited Articles	This article cites 47 articles, 15 of which you can access for free at: http://cancerres.aacrjournals.org/content/70/21/8299.full.html#ref-list-1
Citing Articles	This article has been cited by 1 HighWire-hosted articles. Access the articles at: http://cancerres.aacrjournals.org/content/70/21/8299.full.html#related-urls

E-mail alerts	Sign up to receive free email-alerts related to this article or journal.
Reprints and Subscriptions	To order reprints of this article or to subscribe to the journal, contact the AACR Publications Department at pubs@aacr.org .
Permissions	To request permission to re-use all or part of this article, contact the AACR Publications Department at permissions@aacr.org .

Integrated Microfluidic and Imaging Platform for a Kinase Activity Radioassay to Analyze Minute Patient Cancer Samples

Cong Fang^{1,5}, Yanju Wang^{1,5}, Nam T. Vu^{1,5}, Wei-Yu Lin^{1,5}, Yao-Te Hsieh^{6,7}, Liudmilla Rubbi^{1,5}, Michael E. Phelps^{1,2,3,4,5}, Markus Müschen^{6,7}, Yong-Mi Kim^{6,7}, Arion F. Chatziioannou^{1,3,4,5}, Hsian-Rong Tseng^{1,2,3,4,5}, and Thomas G. Graeber^{1,2,3,4,5}

Abstract

Oncogenic kinase activity and the resulting aberrant growth and survival signaling are a common driving force of cancer. Accordingly, many successful molecularly targeted anticancer therapeutics are directed at inhibiting kinase activity. To assess kinase activity in minute patient samples, we have developed an immunocapture-based *in vitro* kinase assay on an integrated polydimethylsiloxane microfluidics platform that can reproducibly measure kinase activity from as few as 3,000 cells. For this platform, we adopted the standard radiometric ³²P-ATP-labeled phosphate transfer assay. Implementation on a microfluidic device required us to develop methods for repeated trapping and mixing of solid-phase affinity microbeads. We also developed a solid-state beta-particle camera imbedded directly below the microfluidic device for real-time quantitative detection of the signal from this and other microfluidic radiobioassays. We show that the resulting integrated device can measure ABL kinase activity from BCR-ABL-positive leukemia patient samples. The low sample input requirement of the device creates new potential for direct kinase activity experimentation and diagnostics on patient blood, bone marrow, and needle biopsy samples. *Cancer Res*; 70(21); 8299–308. ©2010 AACR.

Introduction

As principal signal transduction components, kinases regulate as many as 50% of intracellular proteins. Aberrant kinase function is involved in the etiology of many diseases and most forms of cancer (1–3). The 518 known human kinases represent one of the largest classes of drug targets pursued by pharmaceutical companies (1, 3), spurred in part by successful targeted inhibition of kinases by antibodies (4) and small-molecule kinase inhibitors (5). Kinase-related signaling measurements directly from patient samples often involve detecting the consequences of activity, for example, the resulting downstream phosphorylation patterns (6–8). Direct measurement of kinase activity from patient samples can

greatly complement these phosphoproteomics techniques (9). There are currently numerous kinase assay technologies with various requirements and ranges of application (10), including a fluorescence-based microfluidic format for screening of kinase inhibition by using purified kinase (11). Radiometric kinase assays represent the earliest technology and are generally considered a high standard for determining basic enzyme properties, in part because radiolabeling of reaction substrates does not alter their intrinsic biochemical and physical properties (10).

Microfluidics offer a prime operation platform for implementation of chemical reactions (12–15) and biological assays (16–18) in a miniaturized fashion due to their inherent advantages of sample and reagent economy, operation fidelity, high throughput, automated operation, and precise control over the microenvironment. The advent of polydimethylsiloxane (PDMS)-based microfluidic devices has enabled the systems-level assembly of individual microfluidic functional modules. These integrated devices provide digitally controlled operation of complicated chemical (19, 20) and biological processes (21, 22) in stand-alone platforms. PDMS devices are fabricated by soft-lithography-based technology (23) in which two layers of fluid and control channels (24) can be designed to create integrated valves and other functional components (15, 19, 21, 22).

Readout directly from microfluidic devices has been accomplished using microelectronic devices coupled to chemical and physical changes in the fluidic space (25) by using an integrated nuclear magnetic resonance sensor (26), and through detection of fluorescent light using avalanche

Authors' Affiliations: ¹Crump Institute for Molecular Imaging, ²Institute for Molecular Medicine, ³Jonsson Comprehensive Cancer Center, ⁴California NanoSystems Institute, and ⁵Department of Molecular and Medical Pharmacology, University of California; ⁶Division of Hematology and Oncology, Childrens Hospital Los Angeles; and ⁷University of Southern California, Los Angeles, California

Note: Supplementary data for this article are available at Cancer Research Online (<http://cancerres.aacrjournals.org/>).

C. Fang and Y. Wang, and H-R. Tseng and T.G. Graeber contributed equally to this work.

Corresponding Authors: Thomas G. Graeber, University of California Los Angeles, 570 Westwood Plaza, CNSI 4341, Los Angeles, CA 90095. Phone: 310-206-6122; Fax: 310-206-8975; E-mail: tgraeber@mednet.ucla.edu or Arion F. Chatziioannou; E-mail: archatziioann@mednet.ucla.edu or Hsian-Rong Tseng; E-mail: hrtseeng@mednet.ucla.edu.

doi: 10.1158/0008-5472.CAN-10-0851

©2010 American Association for Cancer Research.

photodiode detectors (27, 28) or charge-coupled device (CCD) optical camera imaging coupled to microscopy (29) or a scanner (30). However, a wide range of traditional biological and enzymatic assays take advantage of the high detection sensitivity of radioisotopes and use radioactivity as a means to label without altering the chemical properties of the substrates and enzymes. To readout radioactivity distributions directly from a microfluidic chip, we developed a solid-state beta-particle camera by using a position-sensitive avalanche photodiode (PSAPD) charged-particle detector. The beta camera allows real-time quantitative monitoring of the radioassay performance with high sensitivity and low background.

To aid the study of diseases driven by aberrant signaling and the development of kinase inhibition therapies, our goal was to create a miniaturized kinase assay that allows for activity measurements from small patient samples. We therefore developed a microfluidic assay platform that is directly coupled to a beta camera for sensitive and quantitative readout of kinase activity through measurement of ^{32}P incorporation. We developed our microfluidic *in vitro* kinase radioassay (μ -ivkra) by using BCR-ABL oncogenic kinase-positive leukemia samples. Leukemia samples were used because this molecularly defined disease has offered many insights into kinase-driven tumorigenesis (31) and is one of the first successful molecular targets of small-molecule kinase inhibitors, yet still remains a clinical challenge for a subset of cases (32). Our stand-alone benchtop device can measure kinase activity from as few as 3,000 cells and opens new possibilities for experimentation directly on minute patient samples from blood draws, bone marrow aspirates, and needle biopsies.

Materials and Methods

Microfluidic platform fabrication and control

The integrated microfluidic chip was fabricated using an established two-layer soft lithography process and was controlled using pneumatic manifolds digitally instructed through a computer interface (33). Technical details are provided in the Supplementary Data.

Beta camera

The beta camera quantitative radioactivity imaging sensor is composed of a PSAPD silicon semiconductor device with an active area of $14 \times 14 \text{ mm}^2$ (Model P1305-P; RMD Radiation Monitoring Devices), custom electronic readout circuitry, and a computer-based data acquisition card driven by LabVIEW image acquisition software (National Instruments). Charged particles that interact within the depletion region of the PSAPD convert a portion of their kinetic energy into electron-hole pairs. The PSAPD is operated at high-voltage reverse bias (+1,750 V), which accelerates the electron-hole pairs and amplifies the signal by 1,000-fold through an avalanche effect. The PSAPD readout uses a five-channel output with four position channels and one sum channel. The relative amplitudes of the four position channels are used to determine the location of signal along two dimensions. The imaging system also includes a CCD optical camera and reference points that allow for spatial coregistration of the beta camera radio-

activity images with a photographic image of the microfluidic chip. Further details of the beta camera are in the Supplementary Data and will be published elsewhere (N.T.V. and A.F.C.).

Cell culture

Pro-B, lymphoid, Ba/F3 cells transformed with BCR-ABL (p210 isoform) have been described previously and were provided by Charles Sawyers [University of California at Los Angeles (UCLA); ref. 34]. These cells display similar levels of BCR-ABL expression and signaling as patient leukemia primary samples that are positive for the Philadelphia chromosome (Ph+), the chromosomal translocation that results in expression of the BCR-ABL fusion protein. The K562 (Ph+) and U937 (Ph-) human leukemic cell lines were provided by John Colicelli (UCLA). BCR-ABL expression (or lack of expression) was verified by the presence (or absence) of a 210 kDa anti-c-ABL-reactive protein (p210 BCR-ABL isoform; antibody clones K-12, Santa Cruz Biotechnology and OP20, EMD Chemicals) and elevated (or baseline) pan-specific anti-phosphotyrosine levels (clone 4G10, Millipore). Cells were maintained in RPMI 1640 (Cellgro, Mediatech, Inc.) with 10% fetal bovine serum (Omega Scientific). Cells were lysed in modified radioimmunoprecipitation (mRIPA) buffer [10 mmol/L β -glycerophosphate, 50 mmol/L Tris (pH 7.4), 1% NP40, 0.25% Na deoxycolate, 1 mmol/L EDTA, 150 mmol/L NaCl, 1 mmol/L vanadate, with freshly added 1 mmol/L PMSF, 20 $\mu\text{g}/\text{mL}$ leupeptin, 20 $\mu\text{g}/\text{mL}$ aprotinin].

Human leukemia patient samples and mouse xenograft system

Primary cells from the peripheral blood or bone marrow of pre-B-cell acute lymphoblastic leukemia (pre-B-ALL) patients were injected into sublethally irradiated (250 cGy) immunodeficient nonobese/severe combined immunodeficient (NOD/SCID) mice and serially passaged no more than three times (35, 36). The human leukemia cells create a leukemia-like disease burden in the mice and become the dominant subpopulation in the bone marrow, peripheral blood, and spleen. Disease burden was monitored by measuring the percentage of human leukemic cells in the peripheral blood or spleen by using hCD45 flow cytometry. Spleen samples were collected by immediate lysis of scalpel-dissected spleen cells in mRIPA buffer, using the frosted ends of glass slides to promote cell dissociation, on ice. The human cells in the Ph+ samples were uniformly Ph+ based on cytogenetics. The Ph- sample was a pre-B-ALL with a normal karyotype.

Off-chip and on-chip *in vitro* kinase radioassays

In vitro kinase assays were performed using established protocols for bead-based immunoaffinity capture of a specific kinase followed by kinase reactions in the presence of a kinase-specific peptide substrate and ^{32}P -labeled ATP (10, 34). The microfluidic on-chip assay was developed using reagents that were compatible and efficient in the context of microfluidic channels and valves. For the microfluidic assay, 0.1% *n*-dodecyl- β -D-maltoside (DDM) was added to all buffers to prevent bead clumping (indicated by/DDM in the buffer name). DDM (0.1%) had no detectable effect on BCR-ABL

kinase activity (data not shown). For BCR-ABL kinase immunocapture, the off-chip assay used agarose protein A/G beads (40–160 μm ; Pierce) and the on-chip assay used smaller and more sturdy Protein G polystyrene beads (6.7 μm ; Spherotech, Inc.). The large average size and fragility of the agarose beads precluded their use on-chip, and the small size of the polystyrene beads made them impractical for the pipetting-based off-chip assay. Both bead types were coated with anti-c-ABL antibody (OP20, EMD Chemicals). Off-chip assays were performed using 400 μL (20 $\mu\text{g}/\mu\text{L}$, or 4×10^7 cell equivalents) of cell lysate. On-chip assays were performed using the amounts indicated in the figure legends. Abltide-biotin conjugate peptide was used as substrate (Millipore). Following the reaction, the radiolabeled and unlabeled peptides were captured in off-chip assays by using SAM2 Biotin Capture Membrane squares (Promega) or in the on-chip case by using streptavidin-coated polystyrene beads (6.7 μm ; Spherotech, Inc.). Further details for the off-chip and on-chip assays are in the Supplementary Data and Supplementary Fig. S1.

Results

Overall configuration of the microfluidic kinase assay device

To develop a $\mu\text{-ivkra}$ to directly measure kinase activity in minute patient samples, we designed and fabricated a PDMS-based integrated microfluidic chip (24) that performs an immunocapture-based kinase assay (Fig. 1). To enable radioactivity-based readout of the assay, we integrated a PSAPD (37) designed to function as a camera for imaging charged beta particles. The beta camera is imbedded directly below the microfluidic chip, allows real-time monitoring of the radioactivity distribution during the assay, and quantifies the final amount of radioactivity incorporated into the substrate (Fig. 1A). Control and readout of the microfluidic device and beta camera are performed through custom electronics and a personal computer. The resulting device performs an automated multistep kinase reaction assay with coupled readout in a single unit, complete from sample loading to final quantitative data.

Microfluidic device design and operation

The standard *in vitro* kinase radioassay involves immunocapture of the kinase of interest from cell lysate to a solid-phase support, typically using antibody-coated beads. After washing away other kinases and proteins, ^{32}P -ATP and a peptide substrate with specificity to the kinase of interest are incubated with the captured kinase to allow kinase-catalyzed transfer of radiolabeled phosphate from the ATP to the substrate. The peptide substrate (both ^{32}P -labeled and unlabeled) is captured using an affinity resin (e.g., biotin-conjugated peptides captured with a streptavidin solid support membrane) and unincorporated radioisotope (as ^{32}P -ATP) is washed away. To assess kinase activity, the amount of radiolabeled phosphate incorporated into the substrate is quantitatively measured, typically using liquid scintillation.

Our chip design contains two isolated and symmetric fluidic modules for individual kinase reactions run in parallel. The fluidic layer channels of both modules are simultaneously controlled by the pressure actuations of their shared bottom control layer—specifically, each bottom layer control channel regulates a valve in one fluidic module, crosses the midline of the device, and also regulates the corresponding valve in the second module (Fig. 1B; Supplementary Fig. S2). Each module has both (a) an upper circulation chamber and bead-trapping column for the immunocapture and kinase reaction steps and (b) a lower bead-trapping column for the final substrate capture step (Fig. 1C–D). The upper circulation chamber is designed for manipulating the immunocapture beads through multiple trap and release steps (Fig. 2A, steps i–iii).

To miniaturize the kinase radioassay to a microfluidic format, we first developed methods for the reproducible manipulation of solid-support microbeads. In our assay, protein G-conjugated beads are loaded into the chip, coated with anti-ABL antibody, used to capture BCR-ABL kinase from cellular lysates, and then the captured kinase is incubated with ^{32}P -ATP and peptide substrate for the kinase reaction. These steps require three sieve valve-mediated cycles of bead trapping, release, and resuspension for the required solution exchanges and a final retrapping of the immunocapture beads (Supplementary Video 1). In the two-layer integrated PDMS microfluidic platform, we found that polystyrene beads were rigid enough for use in conjunction with the microfluidic valves, and that 6.7- μm -diameter beads could be readily trapped behind sieve-style valves (19) and subsequently released and resuspended homogeneously. When closed, the sieve valves block most of the channel cross-section and thus block bead passage, yet liquid can flow through the remaining opening (further described in Supplementary Methods). For the bead trapping and remixing steps, we optimized our methods to prevent bead clumping and subsequent clogging of the channels. To avert clumping, our final device uses two sets of imbedded peristaltic pump valves per chamber to drive mixing, a pumping protocol with a 2.5-Hz pumping frequency, reversal of the pumping direction every 80 seconds (Supplementary Video 1), and a mild surfactant (0.1% DDM) in the solutions.

At the end of the assay, the kinase reaction mixture is passed through a lower substrate-capture column. This column is preloaded with 6.7- μm streptavidin-coated polystyrene beads to capture the ^{32}P -labeled (and unlabeled) biotin-conjugated peptide substrate while washing away unincorporated radioisotopes (^{32}P -ATP; Fig. 2A, step iv). This single-use column is kept homogeneous by using a flow reversal protocol that promotes efficient capture of the peptide substrate. The flow protocol involves passing the peptide substrate-containing liquid through the bead column in both directions alternatively three times, followed by washing. For this step, the pumping is driven by the noncolumn peristaltic pump in the upper circulation chamber by using a pumping frequency of 5 Hz. The adjoining chip channels are used as reservoirs during the flow reversals. Upon peptide substrate capture in the lower column, the microfluidic steps of the




Figure 1. Schematic of μ -ivkra. **A**, leukemia smear from a BCR-ABL–driven chronic myelogenous leukemia patient (left). In the kinase activity assay, the radiolabeled phosphate from $^{32}\text{P-ATP}$ was transferred to a peptide substrate and kinase activity was quantified based on isotope incorporation (center). Configuration of the μ -ivkra device (right). A microfluidics platform is directly coupled to a highly sensitive charged particle camera (beta camera, PSAPD), enabling on-chip measurements of radioisotope-based kinase assays. **B**, micrograph of a μ -ivkra chip loaded with colored dyes. The microfluidic chip includes two symmetric, isolated reaction units under synchronous digital control. The two-layer microfluidic chip has a top fluidic and a bottom control layer. The colored liquid indicates the delegated functional responsibilities of various components: blue for fluidic channels; in the bottom control layer, red for regular isolation valves, green for sieve valves to trap beads, and yellow for peristaltic pump valves driving fluidic circulation. Scale bar, 2 mm. Additional photographs of the chip platform are in Supplementary Fig. S2. **C**, drawing of the core part of a single unit. **D**, schematic of the core steps in the μ -ivkra: step i, antibody capture using protein G–conjugated polystyrene (PS) beads (6.7- μm diameter); step ii, kinase capture using antibody-coated beads; step iii, kinase reaction; and step iv, biotin-labeled peptide substrate capture by using 6.7- μm -diameter streptavidin-coated polystyrene beads. Steps i to iii are performed in the upper column and circulation chamber where beads can be trapped in the column, released into the circulation chamber (370 nL volume each), mixed to homogeneity, and retrapped multiple times. The sieve valve–mediated bead trapping approach facilitates solution exchange and bead washing (Supplementary Video 1). Capture of substrate (step iv) and washing of unincorporated radioisotope ($^{32}\text{P-ATP}$) occur in the lower column before quantitative radiometric detection imaging using the beta camera.

assay are complete and the ^{32}P incorporated into the peptides is quantitatively measured using the beta camera.

Assay readout through the beta camera

The beta camera serves as the radioactivity readout device for the microfluidic kinase assay. The detector in the beta camera is a $14 \times 14 \text{ mm}^2$ PSAPD with a very low inherent background count rate (1.5 counts/h/ mm^2). The benefits of the PSAPD detector include its monolithic and thus rugged design and position decoding through a five-channel analogue readout. The PSAPD was originally designed for detection of scintillation light produced in nuclear imaging applications (37). We modified the PSAPD with a Mylar pas-

sivation layer to block visible light. When the PSAPD is placed in close contact with a radioactive source, it has a high avalanche-mediated sensitivity for detecting emitted charged particles. Using known levels of ^{32}P , we established a calibration curve and determined the absolute sensitivity of the integrated microfluidic beta camera to be 29% (see Supplementary Fig. S3). The beta camera itself has high intrinsic sensitivity to ^{32}P particles that traverse through the PSAPD detection region. The geometric configuration of the integrated microfluidic beta camera device reduces the overall sensitivity due to (a) half of the beta particles being emitted away from the planar detector and (b) attenuation of beta particles by the control layer of PDMS material (100 μm)

and the glass slide (150 μm) between the microfluidic channel and the beta camera.

The beta camera is position sensitive and thus a single detector is used for multiple simultaneous readouts from adjacent

columns (Fig. 2B–D). In addition to the final readout, the beta camera can be used for qualitative and quantitative operational checkpoints. For example, during the assay, we monitor in real-time the radioactivity distribution and intensity to test for




Figure 2. Schematic diagrams of key steps of the operation of the on-chip kinase radioassay. A, roman numerals correspond to the steps in Fig. 1D. Steps i–ii, loading anti-kinase antibody solution or cell lysate into the circulation chamber (370 nL volume). Releasing protein G beads (6.7- μm diameter, polystyrene) into the circulation chamber for antibody coating [30 min at room temperature (RT); step i] or kinase capture (30 min, RT; step ii); the solution and beads are circulated around the upper chamber driven by two serial sets of peristaltic pump valves. The pumping direction is reversed every 80 s to prevent bead clumping. Retrapping beads in the capture column of the upper circulation chamber between steps (5 min) to allow for bead washing and solution exchange in the circulation chamber. Step iii, kinase reaction with solution and bead circulation in the upper chamber (15 min, RT). The operational steps here are the same as in steps i to ii, except that while waiting for the kinase reaction to finish, the streptavidin polystyrene beads are independently loaded into the lower column (80 nL volume) from below. Step iv, capturing and washing the labeled (and unlabeled) peptide substrate in the lower streptavidin bead substrate-capture column by using a flow reversal protocol before washing away any unincorporated radioisotope (³²P-ATP), and imaging and quantitation by using the beta camera (PSAPD). Different channel colors represent different loaded solutions. B to D, midexperiment monitoring of device and assay performance and final signal quantitation using the beta camera. Overlaid images of the CCD camera optical image, the false-color beta camera radioactivity image, and the chip AutoCAD drawing during various assay steps: loaded radioactive (³²P-ATP) kinase reaction buffer before (B) and after (C) circulation demonstrating equal loading, and (D) the final captured radiolabeled peptide substrate in the lower column showing the anticipated 2-fold difference in kinase activity between 6,000 and 3,000 Ba/F3 + BCR-ABL cells. Color bars, beta camera image scales in counts/s/mm². The total detected activity (pCi) for each region of interest (ROI, yellow outline) is indicated. See Supplementary Methods for how the ROI was chosen.

uniform chip operation during fluid and bead manipulation. These checkpoints verify equal loading, uniform mixing, proper valve operation, and chip channel integrity (Fig. 2B–C; Supplementary Table S1). These checkpoints were useful during the development phase of the microfluidic device and protocol (see the kinase autophosphorylation example below).

On-chip *in vitro* BCR-ABL kinase radioassay

Using the μ -ivkra-integrated microfluidic and beta camera platform described above and in Figs. 1 and 2, we have successfully developed an on-chip *in vitro* BCR-ABL kinase radioassay. The assay adheres to all of the steps of traditional

immunocapture-based *in vitro* kinase radioassays (Fig. 1D). The most notable differences are in terms of scale, including reduced lysate and reagent input, smaller solid-phase bead components, and shorter incubation times (Fig. 2A).

For validation of the microfluidic assay, we performed both a traditional off-chip BCR-ABL kinase radioassay and our on-chip kinase assay by using the same cell lysates (Fig. 3). These assays were done using Ba/F3 cells, a pro-B murine cell line, transfected with either a BCR-ABL expression plasmid or an empty vector control (Fig. 3A). The off-chip kinase assay was performed using millions of cells, whereas the on-chip assay required only 4,500 cells. We observed similar order of




Figure 3. Development of an on-chip *in vitro* BCR-ABL kinase radioassay. A, anti-ABL immunoblot of the Ba/F3 cell line with and without expression of constitutively active BCR-ABL kinase (loading control, anti-actin immunoblotting). B, on-chip BCR-ABL kinase reaction was performed with 12,000 Ba/F3 + BCR-ABL cells (left column) and parental Ba/F3 cells (right column). Final results are represented by an overlay of optical and beta camera false-color images (20-min acquisition) of the lower substrate capture columns. The color bar scale (counts/s/mm²) and total detected activity (pCi) for each ROI (region of interest, yellow circle) are indicated. C, comparison between traditional test tube-based (off-chip) and on-chip *in vitro* BCR-ABL kinase radioassays. Off-chip assays were performed with 10⁷ cells and on-chip assays with 4,500 cells. Results reflect absolute quantitation, and the error bars represent the SD for three assays performed on different days by using aliquots of the same lysate with different chips. D, demonstration of linearity by using BCR-ABL-expressing Ba/F3 cells. In each experiment, samples were diluted with parental non-BCR-ABL-expressing cells as necessary to keep the total cell number constant at 15,000. The graph incorporates data from a total of 16 sample aliquots (eight pair-wise chip experiments performed on different chips using mixed aliquots of the same lysates). The results from each experiment were normalized such that the mean detection signal was set equal to the mean number of cells in the experiment divided by 3,000. Nonspecific background signal was not subtracted. Thus, the results represent relative quantitation designed to identify the linear region of the platform. The R^2 value from a linear regression fit is 0.997. Error bars represent SD.




Figure 4. On-chip *in vitro* BCR-ABL kinase radioassays in leukemic systems. A, detection of BCR-ABL activity from BCR-ABL-expressing human cell lines. The on-chip *in vitro* BCR-ABL kinase radioassay was performed on lysates from the human leukemic cell lines K562 (with leukemia patient-derived BCR-ABL expression) and U937 (with no BCR-ABL expression). B, qualitative detection of BCR-ABL activity from a subpopulation of BCR-ABL-expressing cells as a model for heterogeneous patient samples. BCR-ABL activity was detected from a 1 in 5 subpopulation of BCR-ABL-expressing cells (4,500 Ba/F3 + BCR-ABL: 18,000 Ba/F3 cells versus 22,500 Ba/F3 cells). C, detection of sensitivity to BCR-ABL-targeted drug inhibition. Ba/F3 + BCR-ABL (p210) cells were treated with 125 nmol/L dasatinib or DMSO solvent alone (control) for 2 h, and then BCR-ABL kinase activity was measured in the microfluidic platform (4,500 cell input). D, detection of BCR-ABL activity from Ph⁺ leukemic patient samples in a mouse xenograft system. Mouse spleens with >95% hCD45 Ph⁺ cells or >90% hCD45 Ph⁻ cells were harvested, lysed, and analyzed by anti-c-ABL immunoprecipitation and immunoblotting and anti-TUBB immunoblotting (left) and by the on-chip *in vitro* BCR-ABL kinase radioassay (4,500 cell input; right). All results are represented by overlays of optical images of the lower substrate capture columns and beta camera false-color images (20-min acquisition). The color bar scale (counts/s/mm²) and total detected activity (pCi) for each ROI (region of interest, yellow circle) are indicated. The size of the ROI is uniform across all experiments.

magnitude “BCR-ABL versus control” fold-change values between the off-chip (mean \pm SD, 10.2 ± 1.2 -fold; $n = 3$) and on-chip (14.6 ± 0.7 -fold, $n = 3$) assays (Fig. 3B–C), confirming that the on-chip assay was performed correctly. These absolute quantitation assays were performed using the same lysate split into aliquots, and the low SD in fold change and raw radioactivity signal (Fig. 3C) reflects the reproducibility of the assay performed on different days using different chips. The difference in exact fold change between on- and off-chip results is likely primarily due to differences in the unsubstracted background signal between the two assays and its effect on the accurate measurement of the low ABL activity in the BCR-ABL-free control cells (see further discussion below related to the linearity range of the platform).

We next tested the linearity of the on-chip *in vitro* BCR-ABL kinase radioassay. In these experiments, we loaded different amounts of lysate from BCR-ABL-expressing Ba/F3 cells into each of the microfluidic chip reaction units (with parental Ba/F3 cell lysate used to dilute the amount of BCR-ABL kinase while keeping the total cell number constant). These relative quantitation experiments showed linearity over a 5-fold range of input (Fig. 3D; Supplementary Fig. S4). The lower limit was due mainly to nonspecific background signal from radioisotope bound to the streptavidin beads (the beta camera itself has very low inherent background). Future device designs will use smaller volumes of capture reagents to reduce the background and will incorporate additional reaction units to allow direct determination and subtraction of background to extend this linear range.

We were also able to detect autophosphorylation of BCR-ABL bound to the antibody-coated beads in the upper kinase reaction module at the end of the kinase reaction (Supplementary Fig. S5). During the development of the assay, this readout was used to verify that the reaction had proceeded as anticipated with stronger signal coming from the sample with higher BCR-ABL levels.

The two 30-minute binding steps of our on-chip assay are shorter than those used in conventional immunocapture-based kinase radioassays, which are often reported as from 1 to 4 hours to overnight (10, 38). In total, the on-chip assay can be completed in 4 hours. Only 1 hour elapses from lysate addition to the end of the kinase reaction, thus reducing the potential for decay of kinase activity. This and other advantages of the microfluidic platform, such as efficient, immediate, and consistent mixing, improve the net on-chip kinase activity efficiency by approximately an order of magnitude compared with off-chip assays performed with the same lysates (Supplementary Table S2).

Application of the microfluidic kinase assay on different leukemic systems

We next tested our microfluidic assay on additional BCR-ABL-expressing leukemia systems. We were able to detect substantial BCR-ABL kinase activity differences between patient-derived leukemic cell lines with endogenous BCR-ABL expression (K562) compared with leukemic cell lines without BCR-ABL expression (U937; Fig. 4A).

Because leukemic patient samples are often "contaminated" with normal cells that do not have elevated kinase activity, we tested whether we can measure BCR-ABL kinase activity from a subpopulation of BCR-ABL-expressing cells. We made a 1:4 mixture of BCR-ABL- to non-BCR-ABL-expressing cells by using the Ba/F3 system. The signal from this 20% BCR-ABL-expressing subpopulation was clearly detectable and distinct compared with the background signal from a control non-BCR-ABL-expressing population (Fig. 4B).

One potential application of the on-chip radioactive *in vitro* kinase assay is to measure the inhibition of kinases by molecularly targeted therapeutics directly on patient and animal model samples. To demonstrate this use, we treated BCR-ABL-expressing Ba/F3 cells with 125 nmol/L dasatinib, a well-known BCR-ABL inhibitor (39), and then measured BCR-ABL kinase activity from 4,500 cells. Upon treatment, the BCR-ABL activity was inhibited substantially, with the detected assay signal reduced to background levels (Fig. 4C).

Application of the microfluidic kinase assay on patient samples

Finally, we tested our platform for detecting BCR-ABL kinase activity from clinical samples. We used pre-B-ALL primary patient cells that were both positive and negative for the Ph chromosome and thus also for BCR-ABL protein. We injected these patient cells into NOD/SCID mice in which they proliferate to create a leukemia-like disease burden. Ultimately, the human leukemic cells make up the vast majority of the peripheral blood, bone marrow, and spleen cells (35, 36). Fresh spleens from these leukemia-burdened xeno-

graft mice were processed and loaded onto the microfluidic platform to measure BCR-ABL kinase activity. Signal comparable with our cell line results was obtained from the Ph+ patient samples, whereas the negative control Ph- samples had only background levels of signal (Fig. 4D).

Discussion

By coupling a microfluidic platform to a solid-state beta camera, we have developed an integrated platform for immunocapture-based *in vitro* kinase radioassays on minute cancer samples. The device executes an automated, multi-step solid-phase binding and enzymatic reaction and provides imaging-based final quantitative readout of the assay. In the realm of microfluidics, we have developed new techniques for efficient handling of solid-phase microbeads that allow for the multiple bind, wash, and solvent exchange cycles required by many affinity-based bioassays. In particular, we found that bead trapping, release, and homogenization using sieve valves and peristaltic pumps is an efficient and reproducible approach. Nonetheless, the full development of a clinical device may require further engineering of bead manipulation techniques, or the introduction of redundancy measures, to ensure the robust performance required in a clinical setting.

The integration of the solid-state beta camera directly underneath the microfluidic platform allows imaging the radioisotope distribution both as a final readout and for real-time monitoring of the operational steps of the assay (Fig. 2B–D; Supplementary Table S1). This ability to troubleshoot in real-time proved useful in our assay development and protocol testing phase (Supplementary Fig. S6). The PSAPD-based imaging camera has high sensitivity and very low inherent background (1.5 counts/h/mm²), making feasible the acquisition of very weak signals over time periods longer than the 20 minutes used here. Many bioassays are radioisotope-based; therefore, this sensitive imaging device facilitates the miniaturization of these assays to a microfluidic platform.

There are significant benefits from reducing the radiometric kinase assay from the macroscale to the microscale. First, we reduced the amount of cell input 2 to 3 orders of magnitude compared with conventional and 96-well format assays (38, 40). Similar reduction in the amount of other reagents and radioactivity are another benefit, especially in the context of radiation safety. By porting the assay onto a microfluidic platform, we improved the reproducibility of the assay by introducing digital operation and a well-controlled microenvironment. The microfluidic environment also generally increases net binding and reaction efficiencies (19, 41, 42), as is the case for our assay (Supplementary Table 2). The overall improvements in efficiency and sensitivity expand possibilities for kinase activity measurements in experimentation on and diagnostics of minute patient samples. Future chip designs will increase the number of reaction units per chip to increase the number of samples or kinases that can be assayed simultaneously, and will couple the device to upstream microfluidic modules for efficient sample delivery or automated cell

treatment and lysis (43–45), toward full experimental lab-on-a-chip aspirations (46, 47).

In addition to the assay reported here, there are two other recent advances in high-sensitivity kinase assays that complement one another. One approach involves a phosphorylation-sensitive fluorescent chemosensor readout coupled to an electroconcentration microfluidic device (17). The second approach uses quantitative mass spectrometry as a readout for phosphorylation, allowing for several tens of kinases to be monitored simultaneously by using distinct and kinase-specific peptide substrates (48, 49). As currently developed, both of these assays rely on the peptide sequence for determining kinase specificity and do not benefit from the additional specificity afforded by kinase immunocapture. Both of the microfluidic format assays have the potential to be developed into inexpensive benchtop stand-alone units—for the microfluidic radioassay, this is in part made possible by the compact beta camera device. Disadvantages of the different assays include the need to develop functional chemosensors specific to each kinase for the fluorescent assay, the need of a sophisticated equipment and associated data analysis for the mass spectrometry assay, and the use of radioactivity (although substantially minimized by the microfluidic scale) for our assay. Radio-, fluorescent chemosensor-, and mass spectrometry-based kinase assays will likely continue to be used toward specific applications.

Taken together, the reduced sample input, decreased assay time, and digitally controlled reproducibility of our microfluidic kinase radioassay facilitates direct experimentation on clinical samples that are either precious or perishable. Here,

we show this potential by using a patient sample and a mouse xenograft system. Future experiments will develop reproducible sample collection and measurement conditions for primary patient samples. Other applications include profiling of patient and animal model samples for their kinase inhibitor drug sensitivity, or measurement of kinase activity from stem cells, cancer stem cells, rare immune cells, and other subpopulations, for example, following flow cytometry- or microfluidic-based sorting.

Disclosure of Potential Conflicts of Interest

The University of California and authors C. Fang, Y. Wang, N.T. Vu, A.F. Chatzioannou, H.-R. Tseng, and T.G. Graeber have applied for patents covering the integrated microfluidic beta camera technology and kinase assay.

Acknowledgments

We thank Dr. Jonathan Said (UCLA) for the leukemia smear micrograph and Brian Skaggs for guidance on the BCR-ABL kinase assay.

Grant Support

Jonsson Cancer Center Foundation/UCLA Transdisciplinary Team grant (T.G. Graeber, H.-R. Tseng, and A.F. Chatzioannou). NIH grants R01CA137060 and R01CA139032 (M. Müschen). N.T. Vu is a postdoctoral fellow supported by the National Institute of Biomedical Imaging and Bioengineering (5T32EB002101). Y.M. Kim is supported by the Jean Perkins Foundation and Stop Cancer. T.G. Graeber is an Alfred P. Sloan Research Fellow.

The costs of publication of this article were defrayed in part by the payment of page charges. This article must therefore be hereby marked *advertisement* in accordance with 18 U.S.C. Section 1734 solely to indicate this fact.

Received 04/06/2010; revised 08/04/2010; accepted 08/25/2010; published OnlineFirst 09/13/2010.

References

- Cohen P. The development and therapeutic potential of protein kinase inhibitors. *Curr Opin Chem Biol* 1999;3:459–65.
- Hunter T. Signaling—2000 and beyond. *Cell* 2000;100:113–27.
- Manning G, Whyte DB, Martinez R, Hunter T, Sudarsanam S. The protein kinase complement of the human genome. *Science* 2002; 298:1912–34.
- Pegram MD, Konecny G, Slamon DJ. The molecular and cellular biology of HER2/neu gene amplification/overexpression and the clinical development of Herceptin (trastuzumab) therapy for breast cancer. *Cancer Treat Res* 2000;103:57–75.
- Druker BJ, Sawyers CL, Kantarjian H, et al. Activity of a specific inhibitor of the BCR-ABL tyrosine kinase in the blast crisis of chronic myeloid leukemia and acute lymphoblastic leukemia with the Philadelphia chromosome. *N Engl J Med* 2001;344:1038–42.
- Irish JM, Kotecha N, Nolan GP. Mapping normal and cancer cell signalling networks: towards single-cell proteomics. *Nat Rev Cancer* 2006;6:146–55.
- Gorre ME, Mohammed M, Ellwood K, et al. Clinical resistance to STI-571 cancer therapy caused by BCR-ABL gene mutation or amplification. *Science* 2001;293:876–80.
- Rikova K, Guo A, Zeng Q, et al. Global survey of phosphotyrosine signaling identifies oncogenic kinases in lung cancer. *Cell* 2007; 131:1190–203.
- Janes KA, Albeck JG, Gaudet S, Sorger PK, Lauffenburger DA, Yaffe MB. A systems model of signaling identifies a molecular basis set for cytokine-induced apoptosis. *Science* 2005;310:1646–53.
- Jia Y, Quinn CM, Kwak S, Talanian RV. Current *in vitro* kinase assay technologies: the quest for a universal format. *Curr Drug Discov Technol* 2008;5:59–69.
- Dunne J, Reardon H, Trinh V, Li E, Farinas J. Comparison of on-chip and off-chip microfluidic kinase assay formats. *Assay Drug Dev Technol* 2004;2:121–9.
- Watts P, Haswell SJ. The application of micro reactors for organic synthesis. *Chem Soc Rev* 2005;34:235–46.
- deMello AJ. Control and detection of chemical reactions in microfluidic systems. *Nature* 2006;442:394–402.
- Lin W-Y, Wang Y, Wang S, Tseng H-R. Integrated microfluidic reactors. *Nano Today* 2009;4:470–81.
- Wang J, Sui G, Mocharla VP, et al. Integrated microfluidics for parallel screening of an *in situ* click chemistry library. *Angew Chem Int Ed Engl* 2006;45:5276–81.
- Fan AC, Deb-Basu D, Orban MW, et al. Nanofluidic proteomic assay for serial analysis of oncoprotein activation in clinical specimens. *Nat Med* 2009;15:566–71.
- Lee JH, Cosgrove BD, Lauffenburger DA, Han J. Microfluidic concentration-enhanced cellular kinase activity assay. *J Am Chem Soc* 2009;131:10340–1.
- Sun J, Masterman-Smith M, Graham N, et al. A microfluidic platform for systems pathology: multiparameter single-cell signaling measurements of clinical brain tumor specimens. *Cancer Res* 2010;70:6128–38.
- Lee CC, Sui G, Elizarov A, et al. Multistep synthesis of a radiolabeled imaging probe using integrated microfluidics. *Science* 2005;310: 1793–6.
- Wang Y, Lin WY, Liu K, et al. An integrated microfluidic device for large-scale *in situ* click chemistry screening. *Lab Chip* 2009;9: 2281–5.
- Hansen CL, Skordalakes E, Berger JM, Quake SR. A robust and scalable microfluidic metering method that allows protein crystal growth by free interface diffusion. *Proc Natl Acad Sci U S A* 2002; 99:16531–6.

22. Ottesen EA, Hong JW, Quake SR, Leadbetter JR. Microfluidic digital PCR enables multigene analysis of individual environmental bacteria. *Science* 2006;314:1464–7.
23. Xia YN, Whitesides GM. Soft lithography. *Annu Rev Mater Sci* 1998; 28:153–84.
24. Unger MA, Chou HP, Thorsen T, Scherer A, Quake SR. Monolithic microfabricated valves and pumps by multilayer soft lithography. *Science* 2000;288:113–6.
25. Vykoukal DM, Stone GP, Gascoyne PR, Alt EU, Vykoukal J. Quantitative detection of bioassays with a low-cost image-sensor array for integrated microsystems. *Angew Chem Int Ed Engl* 2009; 48:7649–54.
26. Lee H, Sun E, Ham D, Weissleder R. Chip-NMR biosensor for detection and molecular analysis of cells. *Nat Med* 2008;14:869–74.
27. Kruger J, Singh K, O'Neill A, Jackson C, Morrison A, O'Brien P. Development of a microfluidic device for fluorescence activated cell sorting. *J Micromech Microeng* 2002;12:486–94.
28. Chabinyc ML, Chiu DT, McDonald JC, et al. An integrated fluorescence detection system in poly(dimethylsiloxane) for microfluidic applications. *Anal Chem* 2001;73:4491–8.
29. Chung K, Crane MM, Lu H. Automated on-chip rapid microscopy, phenotyping and sorting of *C. elegans*. *Nat Methods* 2008;5:637–43.
30. Ozawa T, Kinoshita K, Kadowaki S, et al. MAC-CCD system: a novel lymphocyte microwell-array chip system equipped with CCD scanner to generate human monoclonal antibodies against influenza virus. *Lab Chip* 2009;9:158–63.
31. Wong S, Witte ON. The BCR-ABL story: bench to bedside and back. *Annu Rev Immunol* 2004;22:247–306.
32. Ramirez P, DiPersio JF. Therapy options in imatinib failures. *Oncologist* 2008;13:424–34.
33. Duffy DC, McDonald JC, Schueller OJA, Whitesides GM. Rapid prototyping of microfluidic systems in poly(dimethylsiloxane). *Anal Chem* 1998;70:4974–84.
34. Skaggs BJ, Gorre ME, Ryzkin A, et al. Phosphorylation of the ATP-binding loop directs oncogenicity of drug-resistant BCR-ABL mutants. *Proc Natl Acad Sci U S A* 2006;103:19466–71.
35. Trageser D, Iacobucci I, Nahar R, et al. Pre-B cell receptor-mediated cell cycle arrest in Philadelphia chromosome-positive acute lymphoblastic leukemia requires IKAROS function. *J Exp Med* 2009;206: 1739–53.
36. Lock RB, Liem N, Farnsworth ML, et al. The nonobese diabetic/severe combined immunodeficient (NOD/SCID) mouse model of childhood acute lymphoblastic leukemia reveals intrinsic differences in biologic characteristics at diagnosis and relapse. *Blood* 2002;99:4100–8.
37. Shah KS, Farrell R, Grazioso R, Harmon ES, Karplus E. Position-sensitive avalanche photodiodes for γ -ray imaging. *IEEE Trans Nucl Sci* 2002;49:1687–92.
38. Janes KA, Albeck JG, Peng LX, Sorger PK, Lauffenburger DA, Yaffe MB. A high-throughput quantitative multiplex kinase assay for monitoring information flow in signaling networks: application to sepsis-apoptosis. *Mol Cell Proteomics* 2003;2:463–73.
39. Shah NP, Tran C, Lee FY, Chen P, Norris D, Sawyers CL. Overriding imatinib resistance with a novel ABL kinase inhibitor. *Science* 2004; 305:399–401.
40. Wu D, Mand MR, Veach DR, Parker LL, Clarkson B, Kron SJ. A solid-phase Bcr-Abl kinase assay in 96-well hydrogel plates. *Anal Biochem* 2008;375:18–26.
41. Coti KK, Wang Y, Lin WY, et al. A dynamic micromixer for arbitrary control of disguised chemical selectivity. *Chem Commun (Camb)* 2008:3426–8.
42. Zimmermann M, Delamarche E, Wolf M, Hunziker P. Modeling and optimization of high-sensitivity, low-volume microfluidic-based surface immunoassays. *Biomed Microdevices* 2005;7:99–110.
43. Huang S-B, Lee G-B. Pneumatically driven micro-dispenser for sub-micro-liter pipetting. *J Micromech Microeng* 2009; 19:035027.
44. Du WB, Fang Q, He QH, Fang ZL. High-throughput nanoliter sample introduction microfluidic chip-based flow injection analysis system with gravity-driven flows. *Anal Chem* 2005;77:1330–7.
45. Liu K, Xia C, Shen CK-F, van Dam RM. Automated delivery of small fluid volumes through tubing to microfluidic chips. *Proc MicroTAS* 2008:62–4.
46. Hong JW, Quake SR. Integrated nanoliter systems. *Nat Biotechnol* 2003;21:1179–83.
47. El-Ali J, Sorger PK, Jensen KF. Cells on chips. *Nature* 2006;442: 403–11.
48. Kubota K, Anjum R, Yu Y, et al. Sensitive multiplexed analysis of kinase activities and activity-based kinase identification. *Nat Biotechnol* 2009;27:933–40.
49. Yu Y, Anjum R, Kubota K, Rush J, Villen J, Gygi SP. A site-specific, multiplexed kinase activity assay using stable-isotope dilution and high-resolution mass spectrometry. *Proc Natl Acad Sci U S A* 2009;106:11606–11.